

SALIENT CHARACTERISTICS

SIGNAL GENERATOR, CW (250 kHz to 250 MHz)

GE0TK-B

- 1.0 GENERAL DESCRIPTION This procurement requires a solid-state, CW Signal Generator capable of generating a constant-amplitude sine wave with a variable frequency range of 250 kHz to 250 MHz.
- 2.0 CLASSIFICATION The equipment shall meet the requirements of MIL-T-28800D, Type III, Class 5, Style E, Color R for Navy shipboard, submarine, and shore applications with the following modifications and exceptions:
- a. The non-operating temperature requirement is limited to the range of -40°C to +70°C.
  - b. The relative humidity requirement is limited to 95% noncondensating.
  - c. The operating and non-operating altitude requirements are not invoked.
  - d. The EMI requirement is not invoked.
  - e. The warm-up time is extended to one hour.
- 3.0 OPERATIONAL REQUIREMENTS The equipment shall be capable of generating signals within the parameters and accuracies specified herein.
- 3.1 Output Characteristics
- 3.1.1 Frequency
- 3.1.1.1 Range: At least 250 kHz to 250 MHz with a 50 kHz reference frequency output
  - 3.1.1.2 Resolution: At least 3 digit display
  - 3.1.1.3 Accuracy:  $\pm 0.7$  of least significant digit for indicated frequency
- 3.1.2 Amplitude
- 3.1.2.1 Range: At least 5.0 mV to 5.0 Vp-p into a 50 ohm termination
  - 3.1.2.2 Accuracy: At least  $\pm 5\%$  relative to reference frequency
  - 3.1.2.3 Flatness (peak-to-peak, 50 kHz reference): At least  $\pm 1\%$  from 250 kHz to 100 MHz; at least  $\pm 3\%$  from 100 to 250 MHz
- 3.1.3 Harmonics: At least -35 dBc

4.0 GENERAL REQUIREMENTS

4.1 Power: 115/230 Vac  $\pm 10\%$ , 50 to 440 Hz  $\pm 10\%$ , 150 VA maximum

4.2 Dimensions: The total volume of the unit shall not exceed 6,600 cubic cm (400 cubic in ).

4.3 Weight: The overall weight of the unit shall not exceed 5 kg (11 pounds ).

4.4 Calibration Interval: The calibration interval shall be at least 12 months minimum. The equipment shall be within all accuracy requirements specified herein, with a 72% or greater confidence factor following a calibration interval of 12 months.